Associated products

TECHNOZYM® ADAMTS-13 Antigen Control Set

 $2 \times 0.5 \, \text{mL}$



* CELAD	2°C/ .2-8°C20°C. U	
Reference	Presentation	Forma

Control plasmas for the antigenic assay of ADAMTS-13.

Additional high and low quality controls for the TECHNOZYM® ADAMTS-13 antigen ELISA.

Vial



Informations

ADAMTS-13 is an enzyme responsible for the cleavage of von Willebrand factor under blood flow conditions.

TECHNOZYM® ADAMTS-13 Antigen Calibrator Set

TECHNOZYM® ADAMTS-13 Antigen ELISA


A functional defect in the activity of this enzyme leads to the presence of high molecular weight isoforms of vWF, and thus leads to increased platelet aggregation mainly in areas of microvascularization.

This is the known major cause of Thrombotic Thrombocytopenic Purpura (TTP).

Components

- 2 vials x 0.5 mL lyophilized plasma

4-5450663



Characteristics

- Stability 6 months after reconstitution (-20 ° C)
- 2 controls of 0.20 to 0.70 IU / mL. (depending on the lots).

Since October 2015, the controls and calibrators in the kit have been calibrated against the 1st WHO international standard (NIBSC code 12/252) in IU / mL.



